Free Access™ II – Instructions For Use



1. To facilitate patient comfort, apply lip balm or petroleum jelly to the patient's lips prior to inserting Free Access II.

2. Free Access II comes fully assembled. Hold both ends and squeeze the retractors together, gently bending the stainless steel bow. The bow is positioned at the patient's chin and provides necessary tension to hold the device in place without manual assistance. (In some cases, it may be more convenient to position the bow over the bridge of the patient's nose.)

3. Instruct the patient to open his/her mouth. Insert the intraoral extension of each cheek retractor against the inside of the patient's cheeks. Gently guide the patient's lips to fit comfortably in the curve of each lip retractor.

4. The pre-adjusted tension of the stainless steel bow should be comfortable for most patients. However, if an adjustment is necessary, disassemble the device by rotating the plastic interlocking pivot ring at the end of each cheek retractor, make the following adjustments, and reassemble.

a) To reduce tension of the bow, gently bend the bow to increase its curve.

b) To increase tension of the bow, gently bend the bow to gain a straighter curve.

5. To remove Free Access II from the mouth, carefully grasp the left and right cheek retractors and squeeze together to release the tension on the bow, then gently remove the device from the mouth.

Important!
Instructions for Cleaning and Sterilization

General cleaning before sterilization

The product disassembles into three parts for cleaning. Using a mild cleanser and soft brush or sponge, clean and rinse Free Access II of any debris, blood or saliva before sterilization. Free Access II may be autoclaved.

Autoclave

Free Access II may be sterilized by autoclaving at temperatures up to 275°F (135°C) for a minimum of 4 minutes. Sterilization pouch recommended.

Chemical sterilization is not recommended as it will cause discoloration and damage to Free Access II.

To order Free Access™ II call 888-JMORITA (566-7482) or contact your dealer.

J. MORITA USA INC.
Regional Partner

MORITA

J. MORITA USA, INC. | 9 Mason, Irvine, CA 92618 | (949) 581-9600 | info@jmoritausa.com | www.morita.com/usa | L-196 04/18